


PRODUCT TECHNICAL BRIEF

E047 Filament, Part 129108S, for 3G Analyzer Cell

E047 Filament, Part 129108S, for 3G Analyzer Cell - LeakLab solutions for leak testing, vacuum systems and test equipment.

[Open the current source page](#)



Technical Data

PARAMETER	PUBLISHED VALUE / INFORMATION
Product Label	New

Technical and Project Information

Replacement E047 filament assembly, part number 129108S, for a compatible 3G helium leak-detector analyzer cell. Compatibility must be verified from the instrument model, serial number and service documentation before installation.

This service component contains an iridium filament assembly for a compatible helium mass-spectrometer analyzer cell. The source record lists stainless steel, polymer and iridium among its materials.

Do not select an analyzer filament by appearance alone. Confirm detector manufacturer, exact model, analyzer-cell revision, part number 129108S and instrument serial-number range. Installation must follow the service manual's clean-handling, torque, alignment, electrical and vacuum-conditioning procedure.

After replacement, perform the prescribed analyzer checkout, tuning, background stabilization, reference-leak response test and any required calibration. Opening the analyzer can affect warranty, cleanliness and high-voltage safety.

- Helium Reference Leaks
- Scanning Electron Microscopes
- Leak-Detector Accessories Leak-Detector Spare Parts Vacuum-System Filters and Traps
- Leak-Detector Spare Parts
- Vacuum-System Filters and Traps

WORLDWIDE PROJECT CONTACT

Send the test object or system, operating conditions, target sensitivity or pressure range, applicable standard, destination country, quantity and schedule. LeakLab will review the project scope.

mail@leaklab.org

Regarding: E047 Filament, Part 129108S, for 3G Analyzer Cell

Specifications, standards, verification scope, availability and commercial terms must be confirmed for the exact configuration and destination jurisdiction before order or project acceptance.