

## TECHNICAL ARTICLE

# Rules for working with a helium leak detector

Rules for working with a helium leak detector - LeakLab solutions for leak testing, vacuum systems and test equipment.

[Open the current source page](#)

**Ключевые моменты работы с гелиевым течеискателем**

1. Не допускается перемещать включенный прибор;
2. Течеискатель должен быть заземлен перед использованием;
3. Не допускается откачивать течеискателем объекты объемом более 10 литров без вспомогательной системы вакуумирования;
4. Нельзя вакуумировать течеискателем объект более 5 минут, если не удается достичь давления менее 1000 Па;
5. Для смены объекта контроля обязательно остановить режим «Измерение». Отключение изделия от течеискателя допускается только в режиме «Ожидание»;
6. Объект контроля должен быть очищен от загрязнений и воды;
7. Регламентное обслуживание течеискателя выполняется раз в шесть месяцев.

**Leaklab**испытания на герметичность  
mail@leaklab.ru 8(812)715-00-17

## Technical and Project Information

This technical overview covers Rules for working with a helium leak detector. It is retained for technical context; specifications, availability, dates, commercial terms, and compliance statements must be reconfirmed for every current project.

### International project context

For a current application, specifications and availability must be checked against the exact equipment configuration, destination country, operating conditions, and required standards. Legacy announcements do not constitute a current quotation or compliance statement.

### Discuss your application

Send the test object, process conditions, target sensitivity or vacuum range, acceptance criteria, quantity, and schedule. LeakLab will review the information and confirm a suitable technical route.

## WORLDWIDE PROJECT CONTACT

Send the test object or system, operating conditions, target sensitivity or pressure range, applicable standard, destination country, quantity and schedule. LeakLab will review the project scope.

[mail@leaklab.org](mailto:mail@leaklab.org)

Regarding: Rules for working with a helium leak detector

Specifications, standards, verification scope, availability and commercial terms must be confirmed for the exact configuration and destination jurisdiction before order or project acceptance.